

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112822\
 Data File : VN075456.D
 Acq On : 28 Nov 2022 18:11
 Operator : JC\MD
 Sample : VN1128MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1128MBL01

Quant Time: Nov 29 05:33:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	138443	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	239283	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	196305	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	77723	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	39869	53.649	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	107.300%
35) Dibromofluoromethane	8.171	113	47421	50.449	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.900%
50) Toluene-d8	10.565	98	101104	47.756	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.520%
62) 4-Bromofluorobenzene	12.853	95	63251	40.477	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	80.960%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112822\
Data File : VN075456.D
Acq On : 28 Nov 2022 18:11
Operator : JC\MD
Sample : VN1128MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1128MBL01

Quant Time: Nov 29 05:33:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

