

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112823\
 Data File : VN080291.D
 Acq On : 28 Nov 2023 12:42
 Operator : JC\MD
 Sample : VN1128WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1128WBL01

Quant Time: Nov 29 00:33:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	179769	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	332073	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	346357	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	149553	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	159039	53.580	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	107.160%
35) Dibromofluoromethane	8.165	113	129776	52.964	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.920%
50) Toluene-d8	10.565	98	441351	49.968	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.940%
62) 4-Bromofluorobenzene	12.847	95	164158	49.204	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112823\
Data File : VN080291.D
Acq On : 28 Nov 2023 12:42
Operator : JC\MD
Sample : VN1128WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1128WBL01

Quant Time: Nov 29 00:33:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

