

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112823\
 Data File : VN080299.D
 Acq On : 28 Nov 2023 16:15
 Operator : JC\MD
 Sample : 05561-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 A821

Quant Time: Nov 29 00:39:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	178066	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	334380	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	353863	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	153974	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	158088	53.769	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.540%	
35) Dibromofluoromethane	8.165	113	127218	51.562	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.120%	
50) Toluene-d8	10.571	98	442018	49.698	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 99.400%	
62) 4-Bromofluorobenzene	12.853	95	171399	51.020	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 102.040%	
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	6536	6.619	ug/l #	88
20) Methylene Chloride	5.277	84	3594	1.738	ug/l #	92
25) 2-Butanone	7.494	43	2091	1.782	ug/l #	68
29) Tetrahydrofuran	7.853	42	1292	2.051	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112823\
Data File : VN080299.D
Acq On : 28 Nov 2023 16:15
Operator : JC\MD
Sample : 05561-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
A821

Quant Time: Nov 29 00:39:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

