

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112823\
 Data File : VN080300.D
 Acq On : 28 Nov 2023 16:39
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 29 00:39:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	99	0.00
2 T	Dichlorodifluoromethane	0.745	0.624	16.2	83	0.00
3 P	Chloromethane	0.560	0.453	19.1	82	0.00
4 C	Vinyl Chloride	0.593	0.517	12.8#	87	0.00
5 T	Bromomethane	0.402	0.340	15.4	88	0.00
6 T	Chloroethane	0.447	0.350	21.7	94	0.00
7 T	Trichlorofluoromethane	1.236	1.084	12.3	87	0.00
8 T	Diethyl Ether	0.360	0.324	10.0	91	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.565	0.510	9.7	91	0.00
10 T	Methyl Iodide	0.611	0.551	9.8	84	0.00
11 T	Tert butyl alcohol	0.091	0.085	6.6	94	0.00
12 CM	1,1-Dichloroethene	0.528	0.479	9.3#	91	0.00
13 T	Acrolein	0.080	0.081	-1.3	109	0.00
14 T	Allyl chloride	0.665	0.597	10.2	88	0.00
15 T	Acrylonitrile	0.244	0.217	11.1	90	0.00
16 T	Acetone	0.277	0.189	31.8#	78	0.00
17 T	Carbon Disulfide	1.578	1.215	23.0	80	0.00
18 T	Methyl Acetate	0.891	0.797	10.5	97	0.00
19 T	Methyl tert-butyl Ether	1.866	1.762	5.6	92	0.00
20 T	Methylene Chloride	0.581	0.551	5.2	94	0.00
21 T	trans-1,2-Dichloroethene	0.581	0.529	9.0	92	0.00
22 T	Diisopropyl ether	1.471	1.436	2.4	95	0.00
23 T	Vinyl Acetate	0.884	0.796	10.0	86	0.00
24 P	1,1-Dichloroethane	1.071	0.975	9.0	95	0.00
25 T	2-Butanone	0.329	0.254	22.8	83	0.00
26 T	2,2-Dichloropropane	1.063	0.953	10.3	88	0.00
27 T	cis-1,2-Dichloroethene	0.658	0.609	7.4	93	0.00
28 T	Bromochloromethane	0.421	0.391	7.1	99	0.00
29 T	Tetrahydrofuran	0.177	0.160	9.6	88	0.00
30 C	Chloroform	1.198	1.136	5.2#	96	0.00
31 T	Cyclohexane	0.898	0.723	19.5	83	0.00
32 T	1,1,1-Trichloroethane	1.168	1.075	8.0	91	0.00
33 S	1,2-Dichloroethane-d4	0.826	0.794	3.9	103	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	99	0.00
35 S	Dibromofluoromethane	0.369	0.386	-4.6	111	0.00
36 T	1,1-Dichloropropene	0.534	0.496	7.1	90	0.00
37 T	Ethyl Acetate	0.371	0.333	10.2	89	0.00
38 T	Carbon Tetrachloride	0.678	0.628	7.4	90	0.00
39 T	Methylcyclohexane	0.515	0.461	10.5	82	0.00
40 TM	Benzene	1.518	1.410	7.1	94	0.00
41 T	Methacrylonitrile	0.203	0.195	3.9	93	0.00
42 TM	1,2-Dichloroethane	0.621	0.576	7.2	91	0.00
43 T	Isopropyl Acetate	0.721	0.583	19.1	86	0.00
44 TM	Trichloroethene	0.398	0.366	8.0	93	0.00
45 C	1,2-Dichloropropane	0.353	0.341	3.4#	96	0.00
46 T	Dibromomethane	0.267	0.245	8.2	91	0.00
47 T	Bromodichloromethane	0.587	0.568	3.2	95	0.00
48 T	Methyl methacrylate	0.372	0.352	5.4	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112823\
 Data File : VN080300.D
 Acq On : 28 Nov 2023 16:39
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 29 00:39:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.006	0.005	16.7	92	0.00
50 S	Toluene-d8	1.330	1.370	-3.0	104	0.00
51 T	4-Methyl-2-Pentanone	0.373	0.336	9.9	87	0.00
52 CM	Toluene	0.950	0.919	3.3#	94	0.00
53 T	t-1,3-Dichloropropene	0.604	0.571	5.5	90	0.00
54 T	cis-1,3-Dichloropropene	0.617	0.602	2.4	92	0.00
55 T	1,1,2-Trichloroethane	0.363	0.343	5.5	96	0.00
56 T	Ethyl methacrylate	0.370	0.370	0.0	89	0.00
57 T	1,3-Dichloropropane	0.633	0.607	4.1	95	0.00
58 T	2-Chloroethyl Vinyl ether	0.209	0.207	1.0	101	0.00
59 T	2-Hexanone	0.300	0.248	17.3	83	0.00
60 T	Dibromochloromethane	0.434	0.420	3.2	94	0.00
61 T	1,2-Dibromoethane	0.370	0.342	7.6	92	0.00
62 S	4-Bromofluorobenzene	0.502	0.496	1.2	100	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	97	0.00
64 T	Tetrachloroethene	0.382	0.420	-9.9	110	0.00
65 PM	Chlorobenzene	1.133	1.072	5.4	93	0.00
66 T	1,1,1,2-Tetrachloroethane	0.440	0.441	-0.2	94	0.00
67 C	Ethyl Benzene	2.016	1.996	1.0#	93	0.00
68 T	m/p-Xylenes	0.735	0.733	0.3	92	0.00
69 T	o-Xylene	0.693	0.704	-1.6	93	0.00
70 T	Styrene	1.031	1.046	-1.5	92	0.00
71 P	Bromoform	0.326	0.318	2.5	94	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
73 T	Isopropylbenzene	3.885	3.868	0.4	91	0.00
74 T	N-amyl acetate	1.212	1.088	10.2	83	0.00
75 P	1,1,2,2-Tetrachloroethane	1.163	1.011	13.1	89	0.00
76 T	1,2,3-Trichloropropane	1.225	0.997	18.6	80	0.00
77 T	Bromobenzene	0.980	0.935	4.6	92	0.00
78 T	n-propylbenzene	4.345	4.340	0.1	91	0.00
79 T	2-Chlorotoluene	2.822	2.741	2.9	92	0.00
80 T	1,3,5-Trimethylbenzene	3.382	3.341	1.2	91	0.00
81 T	trans-1,4-Dichloro-2-butene	0.363	0.321	11.6	83	0.00
82 T	4-Chlorotoluene	2.891	2.799	3.2	92	0.00
83 T	tert-Butylbenzene	2.712	2.688	0.9	90	0.00
84 T	1,2,4-Trimethylbenzene	3.330	3.354	-0.7	92	0.00
85 T	sec-Butylbenzene	3.431	3.450	-0.6	90	0.00
86 T	p-Isopropyltoluene	3.065	3.121	-1.8	90	0.00
87 T	1,3-Dichlorobenzene	1.765	1.627	7.8	91	0.00
88 T	1,4-Dichlorobenzene	1.774	1.629	8.2	93	0.00
89 T	n-Butylbenzene	2.565	2.458	4.2	84	0.00
90 T	Hexachloroethane	0.598	0.567	5.2	91	0.00
91 T	1,2-Dichlorobenzene	1.689	1.588	6.0	92	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.270	0.234	13.3	88	0.00
93 T	1,2,4-Trichlorobenzene	0.982	0.838	14.7	85	0.00
94 T	Hexachlorobutadiene	0.565	0.469	17.0	86	0.00
95 T	Naphthalene	2.848	2.242	21.3	77	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112823\
Data File : VN080300.D
Acq On : 28 Nov 2023 16:39
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Nov 29 00:39:53 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.921	0.790	14.2	84	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6