

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN112918\
 Data File : VN052567.D
 Acq On : 29 Nov 2018 16:12
 Operator : MD\SY
 Sample : PB115155ZHE#05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB115155ZHE#05

Quant Time: Nov 30 03:22:48 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 28 01:11:38 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1284649	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1949532	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1719228	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	633277	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	664367	47.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.00%	
35) Dibromofluoromethane	7.59	113	604581	50.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.60%	
50) Toluene-d8	10.09	98	2380134	50.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.74%	
62) 4-Bromofluorobenzene	12.40	95	744973	47.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.78%	
Target Compounds						
16) Acetone	3.82	43	32422	8.788	ug/l	97
20) Methylene Chloride	4.56	84	34979	2.423	ug/l	96
43) Isopropyl Acetate	8.17	43	48208	2.535	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN112918\
Data File : VN052567.D
Acq On : 29 Nov 2018 16:12
Operator : MD\SY
Sample : PB115155ZHE#05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB115155ZHE#05

Quant Time: Nov 30 03:22:48 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 28 01:11:38 2018
Response via : Initial Calibration

