

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN112918\
 Data File : VN052571.D
 Acq On : 29 Nov 2018 17:58
 Operator : MD\SY
 Sample : J5768-08
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 OK-01-112818

Quant Time: Nov 30 03:32:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 28 01:11:38 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1281404	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1965908	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1729469	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	664321	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	661160	47.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.78%	
35) Dibromofluoromethane	7.59	113	600012	49.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.02%	
50) Toluene-d8	10.09	98	2390057	50.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.32%	
62) 4-Bromofluorobenzene	12.40	95	758463	47.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.68%	
Target Compounds						
16) Acetone	3.82	43	32686	8.882	ug/l	99
20) Methylene Chloride	4.55	84	67892	4.714	ug/l	97
43) Isopropyl Acetate	8.17	43	53941	2.813	ug/l #	82
95) Naphthalene	15.14	128	20768	4.202	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN112918\
Data File : VN052571.D
Acq On : 29 Nov 2018 17:58
Operator : MD\SY
Sample : J5768-08
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
OK-01-112818

Quant Time: Nov 30 03:32:10 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 28 01:11:38 2018
Response via : Initial Calibration

