

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN112918\
 Data File : VN052574.D
 Acq On : 29 Nov 2018 19:18
 Operator : MD\SY
 Sample : J5767-22
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 CL-02-112818-C

Quant Time: Nov 30 03:45:55 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 28 01:11:38 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1377498	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2104364	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1835987	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	694619	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	710945	47.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.80%	
35) Dibromofluoromethane	7.59	113	644370	49.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.34%	
50) Toluene-d8	10.09	98	2559808	50.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.38%	
62) 4-Bromofluorobenzene	12.40	95	800847	47.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.38%	
Target Compounds						
16) Acetone	3.82	43	32702	8.266	ug/l	93
20) Methylene Chloride	4.55	84	16876915	1090.179	ug/l	99
43) Isopropyl Acetate	8.18	43	51804	2.524	ug/l #	79
95) Naphthalene	15.14	128	19508	4.123	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN112918\
Data File : VN052574.D
Acq On : 29 Nov 2018 19:18
Operator : MD\SY
Sample : J5767-22
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
CL-02-112818-C

Quant Time: Nov 30 03:45:55 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 28 01:11:38 2018
Response via : Initial Calibration

