

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN112918\
 Data File : VN052586.D
 Acq On : 30 Nov 2018 1:30
 Operator : MD\SY
 Sample : J6144-03 10X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 38 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 14B-3

Quant Time: Nov 30 04:22:51 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\624N112618W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Nov 27 00:48:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.20	128	169709	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.59	114	956521	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	787798	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	8.03	65	342127	29.59	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	98.63%
60) 4-Bromofluorobenzene	12.40	95	310931	25.81	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	86.03%
63) Toluene-d8	10.09	98	1110489	30.40	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	101.33%

Target Compounds

					Qvalue
25) Chloroform	7.37	83	1430910	74.611	ug/l 100
41) Bromodichloromethane	9.40	83	142293	10.124	ug/l 97
53) Dibromochloromethane	10.90	129	6740	0.679	ug/l 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN112918\
Data File : VN052586.D
Acq On : 30 Nov 2018 1:30
Operator : MD\SY
Sample : J6144-03 10X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 38 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
14B-3

Quant Time: Nov 30 04:22:51 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\624N112618W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Nov 27 00:48:33 2018
Response via : Initial Calibration

