

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN112918\
 Data File : VN052563.D
 Acq On : 29 Nov 2018 14:25
 Operator : MD\SY
 Sample : PB115155ZHE#02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB115155ZHE#02

Quant Time: Nov 30 03:03:02 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 28 01:11:38 2018
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1326313	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2017400	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1770271	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	670449	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	679005	47.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.04%	
35) Dibromofluoromethane	7.59	113	614146	49.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.76%	
50) Toluene-d8	10.09	98	2445563	50.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.02%	
62) 4-Bromofluorobenzene	12.40	95	780426	47.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.94%	
Target Compounds						
16) Acetone	3.82	43	33299	8.742	ug/l	100
20) Methylene Chloride	4.55	84	33207	2.228	ug/l	96
43) Isopropyl Acetate	8.17	43	47074	2.392	ug/l #	86


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN112918\
Data File : VN052563.D
Acq On : 29 Nov 2018 14:25
Operator : MD\SY
Sample : PB115155ZHE#02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB115155ZHE#02

Quant Time: Nov 30 03:03:02 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 28 01:11:38 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.67 min Scan# 1833

Delta R.T. 0.00 min

Lab File: VN052563.D

Acq: 29 Nov 2018 14:25

Instrument :

MSVOA_N

ClientSampleId :



PB115155ZHE#02

Tgt Ion:168 Resp: 1326313

Ion Ratio Lower Upper



168 100

99 50.6 41.7 62.5



Abundance Ion 167.90 (167.60 to 168.60): VN0

Ion 98.90 (98.60 to 99.60): VN0



#16

Acetone

Concen: 8.742 ug/l

RT: 3.82 min Scan# 638

Delta R.T. 0.01 min

Lab File: VN052563.D



Acq: 29 Nov 2018 14:25

Tgt Ion: 43 Resp: 33299

Ion Ratio Lower Upper


43 100


58 32.7 26.0 39.0



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 58.00 (57.70 to 58.70): VN0





#20

Methylene Chloride

Concen: 2.228 ug/l

RT: 4.55 min Scan# 865

Delta R.T. 0.00 min

Lab File: VN052563.D

Acq: 29 Nov 2018 14:25

Instrument :

MSVOA_N

ClientSampleId :

PB115155ZHE#02

Tgt Ion: 84 Resp: 33207


Ion Ratio Lower Upper

84 100

49 117.5 98.0 147.0

51 39.2 30.1 45.1

86 60.4 50.7 76.1




Abundance Ion 83.90 (83.60 to 84.60): VN0


Ion 48.90 (48.60 to 49.60): VN0

Ion 50.90 (50.60 to 51.60): VN0

Ion 85.90 (85.60 to 86.60): VN0



Time-->



#33

1,2-Dichloroethane-d4

Concen: N.D. ug/l

RT: 8.03 min Scan# 1945

Delta R.T. 0.00 min

Lab File: VN052563.D


Acq: 29 Nov 2018 14:25

Tgt Ion: 65 Resp: 679005

Ion Ratio Lower Upper


65 100

67 54.7 0.0 109.8




Abundance Ion 64.90 (64.60 to 65.60): VN0

Ion 66.90 (66.60 to 67.60): VN0




Time-->






#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.59 min Scan# 2119
 Delta R.T. 0.00 min
 Lab File: VN052563.D
 Acq: 29 Nov 2018 14:25

Instrument :
 MSVOA_N
 ClientSampleId :
 PB115155ZHE#02

Tgt Ion:114 Resp: 2017400
 Ion Ratio Lower Upper
 114 100
 63 20.0 0.0 41.8
 88 15.4 0.0 31.6




Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): VN
 Ion 87.90 (87.60 to 88.60): VN






#35
 Dibromofluoromethane
 Concen: 50.000 ug/l
 RT: 7.59 min Scan# 1810
 Delta R.T. 0.00 min
 Lab File: VN052563.D
 Acq: 29 Nov 2018 14:25

Tgt Ion:113 Resp: 614146
 Ion Ratio Lower Upper
 113 100
 111 101.6 82.0 123.0
 192 22.2 17.0 25.6



Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V





#43

Isopropyl Acetate

Concen: 2.392 ug/l

RT: 8.17 min Scan# 1988

Delta R.T. 0.00 min

Lab File: VN052563.D

Acq: 29 Nov 2018 14:25

Instrument :

MSVOA_N

ClientSampleId :

PB115155ZHE#02


Tgt Ion: 43 Resp: 47074

Ion Ratio Lower Upper

43 100

61 19.7 24.1 36.1#

87 12.1 10.6 15.8




Abundance Ion 43.00 (42.70 to 43.70): VN052470.D (-1968) (-)

Ion 61.00 (60.70 to 61.70): VN052470.D (-1968) (-)

Ion 87.00 (86.70 to 87.70): VN052470.D (-1968) (-)

Time-->

8.10 8.15 8.20




Abundance Ion 43.00 (42.70 to 43.70): VN052563.D (-1894) (-)

Ion 61.00 (60.70 to 61.70): VN052563.D (-1894) (-)

Ion 87.00 (86.70 to 87.70): VN052563.D (-1894) (-)

Time-->

8.10 8.15 8.20



#50

Toluene-d8

Concen: N.D. ug/l

RT: 10.09 min Scan# 2587

Delta R.T. 0.00 min

Lab File: VN052563.D


Acq: 29 Nov 2018 14:25

Tgt Ion: 98 Resp: 2445563

Ion Ratio Lower Upper

98 100

100 64.3 51.3 76.9




Abundance Ion 98.00 (97.70 to 98.70): VN052470.D (-2570) (-)

Ion 100.00 (99.70 to 100.70): VN052470.D (-2570) (-)

Time-->

10.00 10.10 10.20




Abundance Ion 98.00 (97.70 to 98.70): VN052563.D (-2494) (-)

Ion 100.00 (99.70 to 100.70): VN052563.D (-2494) (-)

Time-->

10.00 10.10 10.20



#62

4-Bromofluorobenzene

Concen: N.D. ug/l

RT: 12.40 min Scan# 3306

Delta R.T. 0.00 min

Lab File: VN052563.D

Acq: 29 Nov 2018 14:25

Instrument :

MSVOA_N

ClientSampleId :

PB115155ZHE#02


Tgt Ion: 95 Resp: 780426

Ion Ratio Lower Upper

95 100

174 87.8 0.0 173.0

176 84.8 0.0 166.2




Abundance Ion 94.90 (94.60 to 95.60): VN052470.D (-3291) (-)

Ion 173.80 (173.50 to 174.50): VN052470.D (-3291) (-)

Ion 175.80 (175.50 to 176.50): VN052470.D (-3291) (-)

Time-->




Abundance Ion 94.90 (94.60 to 95.60): VN052563.D (-3212) (-)

Ion 173.80 (173.50 to 174.50): VN052563.D (-3212) (-)

Ion 175.80 (175.50 to 176.50): VN052563.D (-3212) (-)

Time-->



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.41 min Scan# 2997

Delta R.T. 0.00 min

Lab File: VN052563.D

Acq: 29 Nov 2018 14:25


Tgt Ion: 117 Resp: 1770271

Ion Ratio Lower Upper

117 100

82 53.3 43.6 65.4

119 32.0 25.4 38.2




Abundance Ion 116.90 (116.60 to 117.60): VN052470.D (-2980) (-)

Ion 82.00 (81.70 to 82.70): VN052470.D (-2980) (-)

Ion 118.90 (118.60 to 119.60): VN052470.D (-2980) (-)

Time-->




Abundance Ion 116.90 (116.60 to 117.60): VN052563.D (-2904) (-)

Ion 82.00 (81.70 to 82.70): VN052563.D (-2904) (-)

Ion 118.90 (118.60 to 119.60): VN052563.D (-2904) (-)

Time-->



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.35 min Scan# 3599

Delta R.T. 0.00 min

Lab File: VN052563.D


Acq: 29 Nov 2018 14:25

Instrument :

MSVOA_N

ClientSampleId :

PB115155ZHE#02




Tgt Ion:152 Resp: 670449

Ion Ratio Lower Upper

152 100

115 56.9 28.4 85.2

150 157.5 0.0 346.0



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

