

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN112918\
 Data File : VN052574.D
 Acq On : 29 Nov 2018 19:18
 Operator : MD\SY
 Sample : J5767-22
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 CL-02-112818-C

Quant Time: Nov 30 03:45:55 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 28 01:11:38 2018
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1377498	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2104364	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1835987	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	694619	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	710945	47.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.80%	
35) Dibromofluoromethane	7.59	113	644370	49.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.34%	
50) Toluene-d8	10.09	98	2559808	50.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.38%	
62) 4-Bromofluorobenzene	12.40	95	800847	47.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.38%	
Target Compounds						
16) Acetone	3.82	43	32702	8.266	ug/l	93
20) Methylene Chloride	4.55	84	16876915	1090.179	ug/l	99
43) Isopropyl Acetate	8.18	43	51804	2.524	ug/l #	79
95) Naphthalene	15.14	128	19508	4.123	ug/l	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN112918\
Data File : VN052574.D
Acq On : 29 Nov 2018 19:18
Operator : MD\SY
Sample : J5767-22
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
CL-02-112818-C

Quant Time: Nov 30 03:45:55 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 28 01:11:38 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.67 min Scan# 1833

Delta R.T. 0.00 min

Lab File: VN052574.D

Acq: 29 Nov 2018 19:18

Instrument :

MSVOA_N

ClientSampleId :





CL-02-112818-C

Tgt Ion:168 Resp: 1377498

Ion Ratio Lower Upper

168 100

99 50.3 41.7 62.5



#16

Acetone

Concen: 8.266 ug/l

RT: 3.82 min Scan# 637

Delta R.T. 0.01 min

Lab File: VN052574.D




Acq: 29 Nov 2018 19:18


Tgt Ion: 43 Resp: 32702

Ion Ratio Lower Upper

43 100

58 28.5 26.0 39.0





#20

Methylene Chloride

Concen: 1090.179 ug/l

RT: 4.55 min Scan# 865

Delta R.T. 0.00 min

Lab File: VN052574.D

Acq: 29 Nov 2018 19:18

Instrument :

MSVOA_N

ClientSampleId :

CL-02-112818-C

Tgt Ion: 84 Resp:16876915


Ion Ratio Lower Upper

84 100

49 120.9 98.0 147.0

51 37.8 30.1 45.1

86 64.2 50.7 76.1






Abundance Ion 83.90 (83.60 to 84.60): VN0

Ion 48.90 (48.60 to 49.60): VN0

Ion 50.90 (50.60 to 51.60): VN0

Ion 85.90 (85.60 to 86.60): VN0



#33

1,2-Dichloroethane-d4

Concen: Below ug/l

RT: 8.03 min Scan# 1945

Delta R.T. 0.00 min

Lab File: VN052574.D


Acq: 29 Nov 2018 19:18

Tgt Ion: 65 Resp: 710945

Ion Ratio Lower Upper



65 100


67 54.7 0.0 109.8



Abundance Ion 64.90 (64.60 to 65.60): VN0

Ion 66.90 (66.60 to 67.60): VN0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.59 min Scan# 2119

Delta R.T. 0.00 min

Lab File: VN052574.D

Acq: 29 Nov 2018 19:18

Instrument :

MSVOA_N

ClientSampleId :

CL-02-112818-C


Tgt Ion:114 Resp: 2104364

Ion Ratio Lower Upper

114 100

63 20.0 0.0 41.8




88 15.5 0.0 31.6



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VN0

Ion 87.90 (87.60 to 88.60): VN0



#35

Dibromofluoromethane

Concen: 50.000 ug/l

RT: 7.59 min Scan# 1810

Delta R.T. 0.00 min

Lab File: VN052574.D

Acq: 29 Nov 2018 19:18


Tgt Ion:113 Resp: 644370

Ion Ratio Lower Upper

113 100

111 103.1 82.0 123.0



192 22.2 17.0 25.6




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#43

Isopropyl Acetate

Concen: 2.524 ug/l

RT: 8.18 min Scan# 1991

Delta R.T. 0.01 min

Lab File: VN052574.D

Acq: 29 Nov 2018 19:18

Instrument :

MSVOA_N

ClientSampleId :

CL-02-112818-C


Tgt Ion: 43 Resp: 51804

Ion Ratio Lower Upper

43 100

61 15.8 24.1 36.1#




87 9.6 10.6 15.8#



Abundance Ion 43.00 (42.70 to 43.70): VN052470.D (-1968) (-)

Ion 61.00 (60.70 to 61.70): VN052470.D (-1968) (-)

Ion 87.00 (86.70 to 87.70): VN052470.D (-1968) (-)



#50

Toluene-d8

Concen: N.D. ug/l

RT: 10.09 min Scan# 2587

Delta R.T. 0.00 min

Lab File: VN052574.D


Acq: 29 Nov 2018 19:18

Tgt Ion: 98 Resp: 2559808

Ion Ratio Lower Upper



98 100


100 63.9 51.3 76.9



Abundance Ion 98.00 (97.70 to 98.70): VN052470.D (-2570) (-)

Ion 100.00 (99.70 to 100.70): VN052470.D (-2570) (-)





#62

4-Bromofluorobenzene

Concen: N.D. ug/l

RT: 12.40 min Scan# 3306

Delta R.T. 0.00 min

Lab File: VN052574.D

Acq: 29 Nov 2018 19:18

Instrument :

MSVOA_N

ClientSampleId :

CL-02-112818-C





Tgt Ion: 95 Resp: 800847

Ion Ratio Lower Upper

95 100

174 88.5 0.0 173.0

176 86.0 0.0 166.2



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.41 min Scan# 2997

Delta R.T. 0.00 min

Lab File: VN052574.D

Acq: 29 Nov 2018 19:18




Tgt Ion: 117 Resp: 1835987


Ion Ratio Lower Upper

117 100

82 53.9 43.6 65.4

119 32.1 25.4 38.2





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.35 min Scan# 3599

Delta R.T. 0.00 min

Lab File: VN052574.D

Acq: 29 Nov 2018 19:18

Instrument :

MSVOA_N

ClientSampleId :

CL-02-112818-C


Tgt Ion:152 Resp: 694619

Ion Ratio Lower Upper

152 100

115 56.2 28.4 85.2




150 156.3 0.0 346.0



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V



#95

Naphthalene

Concen: 4.123 ug/l

RT: 15.14 min Scan# 4156

Delta R.T. 0.00 min

Lab File: VN052574.D

Acq: 29 Nov 2018 19:18


Tgt Ion:128 Resp: 19508

Ion Ratio Lower Upper

128 100

127 12.6 10.2 15.2

129 12.2 8.7 13.1



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V

