


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN112919\
 Data File : VN059435.D
 Acq On : 29 Nov 2019 8:56
 Operator : JC/SP
 Sample : BFB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112719W.M
 Title : SW846 8260
 Last Update : Wed Nov 27 16:56:05 2019



AutoFind: Scans 3304, 3305, 3306; Background Corrected with Scan 3291

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.4	7237	PASS
75	95	30	60	47.4	19736	PASS
95	95	100	100	100.0	41613	PASS
96	95	5	9	7.1	2942	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	78.3	32592	PASS
175	174	5	9	7.9	2585	PASS
176	174	95	101	95.4	31090	PASS
177	176	5	9	6.9	2160	PASS