

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112921\
 Data File : VN069660.D
 Acq On : 29 Nov 2021 12:46
 Operator : JC/MD
 Sample : VN1129WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1129WBL01

Quant Time: Nov 30 04:16:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 09 18:18:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1582371	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2747709	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2781118	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1026399	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	1167360	50.880	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	101.760%
35) Dibromofluoromethane	8.024	113	805200	48.860	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.720%
50) Toluene-d8	10.440	98	3495597	53.933	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.860%
62) 4-Bromofluorobenzene	12.731	95	1257877	52.829	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	105.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112921\
Data File : VN069660.D
Acq On : 29 Nov 2021 12:46
Operator : JC/MD
Sample : VN1129WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1129WBL01

Quant Time: Nov 30 04:16:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 09 18:18:22 2021
Response via : Initial Calibration

