

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112921\
 Data File : VN069673.D
 Acq On : 29 Nov 2021 19:09
 Operator : JC/MD
 Sample : M4852-03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-13

Quant Time: Nov 30 04:19:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 09 18:18:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1564347	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2695704	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2780659	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1274152	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	1137430	50.146	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	= 100.300%		
35) Dibromofluoromethane	8.024	113	786889	48.670	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	= 97.340%		
50) Toluene-d8	10.440	98	3495714	54.975	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	= 109.960%		
62) 4-Bromofluorobenzene	12.731	95	1376534	58.928	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	= 117.860%		
Target Compounds						
					Qvalue	
16) Acetone	4.299	43	15586	1.274	ug/l	94
27) cis-1,2-Dichloroethene	7.332	96	24736	1.409	ug/l	96
39) Methylcyclohexane	9.461	83	1129858	40.042	ug/l	90
40) Benzene	8.461	78	646299	9.231	ug/l	100
52) Toluene	10.507	92	15476	0.355	ug/l	88
67) Ethyl Benzene	11.843	91	320146	3.354	ug/l	97
68) m/p-Xylenes	11.953	106	65927	1.878	ug/l	82
69) o-Xylene	12.280	106	28024	0.801	ug/l	59
73) Isopropylbenzene	12.578	105	5642153	55.615	ug/l	99
78) n-propylbenzene	12.921	91	8078505	70.769	ug/l	96
80) 1,3,5-Trimethylbenzene	13.058	105	1265402	15.429	ug/l	98
83) tert-Butylbenzene	13.323	119	742218	10.327	ug/l	92
84) 1,2,4-Trimethylbenzene	13.366	105	4039683	51.022	ug/l	98
85) sec-Butylbenzene	13.500	105	2392578	23.831	ug/l	81
95) Naphthalene	15.504	128	1277366	25.024	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112921\
Data File : VN069673.D
Acq On : 29 Nov 2021 19:09
Operator : JC/MD
Sample : M4852-03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-13

Quant Time: Nov 30 04:19:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 09 18:18:22 2021
Response via : Initial Calibration

