

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112922\
 Data File : VN075464.D
 Acq On : 29 Nov 2022 10:09
 Operator : JC\MD
 Sample : VN1129MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1129MBL01

Quant Time: Nov 30 02:24:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	174313	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	284075	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	232718	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	96302	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	25786	27.558	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	55.120%#
35) Dibromofluoromethane	8.165	113	30323	27.173	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	54.340%#
50) Toluene-d8	10.565	98	65704	26.141	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	52.280%#
62) 4-Bromofluorobenzene	12.853	95	41627	22.438	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	44.880%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112922\
Data File : VN075464.D
Acq On : 29 Nov 2022 10:09
Operator : JC\MD
Sample : VN1129MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1129MBL01

Quant Time: Nov 30 02:24:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

