

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112922\
 Data File : VN075472.D
 Acq On : 29 Nov 2022 13:39
 Operator : JC\MD
 Sample : N5793-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-2022-11-28

Quant Time: Nov 30 02:27:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	122877	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	211247	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	176718	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	69752	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	34013	51.567	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.140%
35) Dibromofluoromethane	8.171	113	40817	49.186	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.380%
50) Toluene-d8	10.571	98	81411	43.557	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	87.120%
62) 4-Bromofluorobenzene	12.847	95	59980	43.478	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.960%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112922\
Data File : VN075472.D
Acq On : 29 Nov 2022 13:39
Operator : JC\MD
Sample : N5793-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-2022-11-28

Quant Time: Nov 30 02:27:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

