

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112922\
 Data File : VN075475.D
 Acq On : 29 Nov 2022 14:51
 Operator : JC\MD
 Sample : PB149314ZHE#01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB149314ZHE#01

Quant Time: Nov 30 02:27:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	127498	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	224377	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	206810	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	80316	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	39811	58.170	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	116.340%
35) Dibromofluoromethane	8.171	113	48542	55.072	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	110.140%
50) Toluene-d8	10.571	98	104156	52.466	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.940%
62) 4-Bromofluorobenzene	12.847	95	70951	48.421	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.840%
Target Compounds						
					Qvalue	
16) Acetone	4.448	43	12061	18.893	ug/l	91
20) Methylene Chloride	5.295	84	7723	4.810	ug/l #	71
37) Ethyl Acetate	7.571	43	17235	9.958	ug/l	98
43) Isopropyl Acetate	8.694	43	104002	37.436	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112922\
Data File : VN075475.D
Acq On : 29 Nov 2022 14:51
Operator : JC\MD
Sample : PB149314ZHE#01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB149314ZHE#01

Quant Time: Nov 30 02:27:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

