

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112922\
 Data File : VN075481.D
 Acq On : 29 Nov 2022 17:14
 Operator : JC\MD
 Sample : N5780-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-M

Quant Time: Nov 30 02:29:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	127448	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	225448	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	205461	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	80034	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	41028	59.972	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 119.940%		
35) Dibromofluoromethane	8.171	113	50215	56.700	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 113.400%		
50) Toluene-d8	10.565	98	109158	54.724	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 109.440%		
62) 4-Bromofluorobenzene	12.847	95	78610	53.393	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 106.780%		
Target Compounds						
					Qvalue	
16) Acetone	4.448	43	12181	19.089	ug/l	92
20) Methylene Chloride	5.283	84	7672	4.772	ug/l #	87
37) Ethyl Acetate	7.571	43	16713	9.610	ug/l	97
43) Isopropyl Acetate	8.688	43	122492	43.882	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112922\
Data File : VN075481.D
Acq On : 29 Nov 2022 17:14
Operator : JC\MD
Sample : N5780-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-M

Quant Time: Nov 30 02:29:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

