

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112922\
 Data File : VN075483.D
 Acq On : 29 Nov 2022 18:02
 Operator : JC\MD
 Sample : N5795-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-10

Quant Time: Nov 30 02:30:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	141734	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	248179	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.864	117	225935	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	90431	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	34030	44.729	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	89.460%
35) Dibromofluoromethane	8.171	113	39999	41.028	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	82.060%
50) Toluene-d8	10.570	98	89934	40.957	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	81.920%#
62) 4-Bromofluorobenzene	12.847	95	65693	40.533	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	81.060%#
Target Compounds						
					Qvalue	
16) Acetone	4.453	43	12600	17.755	ug/l	99
18) Methyl Acetate	5.047	43	4666	1.348	ug/l	96
20) Methylene Chloride	5.294	84	5833	2.847	ug/l #	84
37) Ethyl Acetate	7.565	43	17694	9.243	ug/l	98
43) Isopropyl Acetate	8.694	43	134885	43.896	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112922\
Data File : VN075483.D
Acq On : 29 Nov 2022 18:02
Operator : JC\MD
Sample : N5795-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-10

Quant Time: Nov 30 02:30:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

