

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112922\
 Data File : VN075484.D
 Acq On : 29 Nov 2022 18:26
 Operator : JC\MD
 Sample : N5795-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-12

Manual Integrations
 APPROVED

Reviewed By : John Carlone 11/30/2022
 Supervised By : Mahesh Dadoda 11/30/2022

Quant Time: Nov 30 02:30:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	127129	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	229644	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	207354	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	78037	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	34065	49.919	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	99.840%		
35) Dibromofluoromethane	8.177	113	42164	46.739	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	93.480%		
50) Toluene-d8	10.571	98	86725	42.683	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	85.360%#		
62) 4-Bromofluorobenzene	12.853	95	63186	42.132	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	84.260%		
Target Compounds						
16) Acetone	4.442	43	14061	22.090	ug/l	91
18) Methyl Acetate	5.047	43	4288	1.414	ug/l	92
20) Methylene Chloride	5.271	84	6217m	3.631	ug/l	
37) Ethyl Acetate	7.565	43	16692	9.423	ug/l	99
43) Isopropyl Acetate	8.694	43	101781	35.796	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112922\
Data File : VN075484.D
Acq On : 29 Nov 2022 18:26
Operator : JC\MD
Sample : N5795-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-12

Quant Time: Nov 30 02:30:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 11/30/2022
Supervised By : Mahesh Dadoda 11/30/2022

