

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112923\
 Data File : VN080307.D
 Acq On : 29 Nov 2023 14:10
 Operator : JC\MD
 Sample : 05470-01
 Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-1

Quant Time: Nov 30 04:44:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	167905	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	303003	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	308299	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	141196	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	145946	52.643	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.280%
35) Dibromofluoromethane	8.159	113	116120	51.938	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.880%
50) Toluene-d8	10.565	98	400550	49.699	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.400%
62) 4-Bromofluorobenzene	12.853	95	150198	49.339	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.680%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112923\
Data File : VN080307.D
Acq On : 29 Nov 2023 14:10
Operator : JC\MD
Sample : 05470-01
Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SB-1

Quant Time: Nov 30 04:44:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

