

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112923\
 Data File : VN080302.D
 Acq On : 29 Nov 2023 11:11
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 30 04:41:31 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	87	0.00
2 T	Dichlorodifluoromethane	0.745	0.678	9.0	79	0.00
3 P	Chloromethane	0.560	0.502	10.4	80	0.00
4 C	Vinyl Chloride	0.593	0.557	6.1#	83	0.00
5 T	Bromomethane	0.402	0.378	6.0	87	0.00
6 T	Chloroethane	0.447	0.420	6.0	99	0.00
7 T	Trichlorofluoromethane	1.236	1.150	7.0	81	0.00
8 T	Diethyl Ether	0.360	0.342	5.0	84	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.565	0.525	7.1	83	0.00
10 T	Methyl Iodide	0.611	0.578	5.4	77	0.00
11 T	Tert butyl alcohol	0.091	0.074	18.7	71	-0.01
12 CM	1,1-Dichloroethene	0.528	0.506	4.2#	84	0.00
13 T	Acrolein	0.080	0.096	-20.0	113	0.00
14 T	Allyl chloride	0.665	0.616	7.4	80	0.00
15 T	Acrylonitrile	0.244	0.216	11.5	79	0.00
16 T	Acetone	0.277	0.248	10.5	90	0.00
17 T	Carbon Disulfide	1.578	1.287	18.4	75	0.00
18 T	Methyl Acetate	0.891	0.807	9.4	86	0.00
19 T	Methyl tert-butyl Ether	1.866	1.848	1.0	85	0.00
20 T	Methylene Chloride	0.581	0.586	-0.9	88	0.00
21 T	trans-1,2-Dichloroethene	0.581	0.556	4.3	85	0.00
22 T	Diisopropyl ether	1.471	1.510	-2.7	88	0.00
23 T	Vinyl Acetate	0.884	0.817	7.6	78	0.00
24 P	1,1-Dichloroethane	1.071	1.012	5.5	87	0.00
25 T	2-Butanone	0.329	0.273	17.0	78	0.00
26 T	2,2-Dichloropropane	1.063	1.016	4.4	83	0.00
27 T	cis-1,2-Dichloroethene	0.658	0.635	3.5	86	0.00
28 T	Bromochloromethane	0.421	0.409	2.9	91	0.00
29 T	Tetrahydrofuran	0.177	0.150	15.3	72	0.00
30 C	Chloroform	1.198	1.203	-0.4#	89	0.00
31 T	Cyclohexane	0.898	0.734	18.3	74	0.00
32 T	1,1,1-Trichloroethane	1.168	1.143	2.1	85	0.00
33 S	1,2-Dichloroethane-d4	0.826	0.813	1.6	93	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	85	0.00
35 S	Dibromofluoromethane	0.369	0.391	-6.0	97	0.00
36 T	1,1-Dichloropropene	0.534	0.522	2.2	82	0.00
37 T	Ethyl Acetate	0.371	0.326	12.1	76	0.00
38 T	Carbon Tetrachloride	0.678	0.687	-1.3	85	0.00
39 T	Methylcyclohexane	0.515	0.488	5.2	75	0.00
40 TM	Benzene	1.518	1.497	1.4	86	0.00
41 T	Methacrylonitrile	0.203	0.187	7.9	77	0.00
42 TM	1,2-Dichloroethane	0.621	0.635	-2.3	87	0.00
43 T	Isopropyl Acetate	0.721	0.580	19.6	74	0.00
44 TM	Trichloroethene	0.398	0.401	-0.8	88	0.00
45 C	1,2-Dichloropropane	0.353	0.363	-2.8#	88	0.00
46 T	Dibromomethane	0.267	0.270	-1.1	87	0.00
47 T	Bromodichloromethane	0.587	0.629	-7.2	90	0.00
48 T	Methyl methacrylate	0.372	0.361	3.0	80	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112923\
 Data File : VN080302.D
 Acq On : 29 Nov 2023 11:11
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 30 04:41:31 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.006	0.005	16.7	74	0.00
50 S	Toluene-d8	1.330	1.404	-5.6	92	0.00
51 T	4-Methyl-2-Pentanone	0.373	0.338	9.4	76	0.00
52 CM	Toluene	0.950	0.965	-1.6#	85	0.00
53 T	t-1,3-Dichloropropene	0.604	0.614	-1.7	84	0.00
54 T	cis-1,3-Dichloropropene	0.617	0.660	-7.0	87	0.00
55 T	1,1,2-Trichloroethane	0.363	0.376	-3.6	91	0.00
56 T	Ethyl methacrylate	0.370	0.392	-5.9	81	0.00
57 T	1,3-Dichloropropane	0.633	0.644	-1.7	87	0.00
58 T	2-Chloroethyl Vinyl ether	0.209	0.192	8.1	80	0.00
59 T	2-Hexanone	0.300	0.260	13.3	75	0.00
60 T	Dibromochloromethane	0.434	0.467	-7.6	90	0.00
61 T	1,2-Dibromoethane	0.370	0.369	0.3	86	0.00
62 S	4-Bromofluorobenzene	0.502	0.547	-9.0	95	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	88	0.00
64 T	Tetrachloroethene	0.382	0.429	-12.3	102	0.00
65 PM	Chlorobenzene	1.133	1.101	2.8	87	0.00
66 T	1,1,1,2-Tetrachloroethane	0.440	0.465	-5.7	90	0.00
67 C	Ethyl Benzene	2.016	2.055	-1.9#	87	0.00
68 T	m/p-Xylenes	0.735	0.759	-3.3	86	0.00
69 T	o-Xylene	0.693	0.729	-5.2	87	0.00
70 T	Styrene	1.031	1.115	-8.1	89	0.00
71 P	Bromoform	0.326	0.333	-2.1	89	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	0.00
73 T	Isopropylbenzene	3.885	3.864	0.5	86	0.00
74 T	N-amyl acetate	1.212	0.994	18.0	71	0.00
75 P	1,1,2,2-Tetrachloroethane	1.163	0.965	17.0	79	0.00
76 T	1,2,3-Trichloropropane	1.225	1.063	13.2	80	0.00
77 T	Bromobenzene	0.980	0.957	2.3	89	0.00
78 T	n-propylbenzene	4.345	4.280	1.5	84	0.00
79 T	2-Chlorotoluene	2.822	2.790	1.1	87	0.00
80 T	1,3,5-Trimethylbenzene	3.382	3.381	0.0	87	0.00
81 T	trans-1,4-Dichloro-2-butene	0.363	0.316	12.9	76	0.00
82 T	4-Chlorotoluene	2.891	2.852	1.3	88	0.00
83 T	tert-Butylbenzene	2.712	2.665	1.7	83	0.00
84 T	1,2,4-Trimethylbenzene	3.330	3.390	-1.8	87	0.00
85 T	sec-Butylbenzene	3.431	3.451	-0.6	84	0.00
86 T	p-Isopropyltoluene	3.065	3.175	-3.6	86	0.00
87 T	1,3-Dichlorobenzene	1.765	1.657	6.1	87	0.00
88 T	1,4-Dichlorobenzene	1.774	1.658	6.5	88	0.00
89 T	n-Butylbenzene	2.565	2.592	-1.1	83	0.00
90 T	Hexachloroethane	0.598	0.561	6.2	84	0.00
91 T	1,2-Dichlorobenzene	1.689	1.616	4.3	88	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.270	0.216	20.0	76	0.00
93 T	1,2,4-Trichlorobenzene	0.982	0.890	9.4	84	0.00
94 T	Hexachlorobutadiene	0.565	0.518	8.3	89	0.00
95 T	Naphthalene	2.848	2.245	21.2	72	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112923\
Data File : VN080302.D
Acq On : 29 Nov 2023 11:11
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Nov 30 04:41:31 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.921	0.831	9.8	83	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6