

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN113018\
 Data File : VN052606.D
 Acq On : 30 Nov 2018 14:54
 Operator : MD\SY
 Sample : PB115186ZHE#04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB115186ZHE#04

Quant Time: Dec 01 02:14:24 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 28 01:11:38 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1200101	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1843243	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1623408	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	588686	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	632601	48.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.82%	
35) Dibromofluoromethane	7.59	113	573280	50.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.90%	
50) Toluene-d8	10.09	98	2254541	50.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.94%	
62) 4-Bromofluorobenzene	12.40	95	702434	47.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.52%	
Target Compounds						
16) Acetone	3.82	43	23273	6.753	ug/l	98
20) Methylene Chloride	4.56	84	18310	1.358	ug/l	98
43) Isopropyl Acetate	8.17	43	61024	3.395	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN113018\
Data File : VN052606.D
Acq On : 30 Nov 2018 14:54
Operator : MD\SY
Sample : PB115186ZHE#04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB115186ZHE#04

Quant Time: Dec 01 02:14:24 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 28 01:11:38 2018
Response via : Initial Calibration

