

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN113018\
 Data File : VN052607.D
 Acq On : 30 Nov 2018 15:21
 Operator : MD\SY
 Sample : PB115186ZHE#05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB115186ZHE#05

Quant Time: Dec 01 02:16:11 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 28 01:11:38 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1226606	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1897886	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1677654	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	653705	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	646376	48.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.80%	
35) Dibromofluoromethane	7.59	113	577383	49.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.70%	
50) Toluene-d8	10.09	98	2308324	50.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.36%	
62) 4-Bromofluorobenzene	12.40	95	744623	48.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.30%	
Target Compounds						
16) Acetone	3.82	43	24349	6.912	ug/l	99
20) Methylene Chloride	4.55	84	21097	1.530	ug/l	95
43) Isopropyl Acetate	8.17	43	68352	3.693	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN113018\
Data File : VN052607.D
Acq On : 30 Nov 2018 15:21
Operator : MD\SY
Sample : PB115186ZHE#05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB115186ZHE#05

Quant Time: Dec 01 02:16:11 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 28 01:11:38 2018
Response via : Initial Calibration

