

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN113018\
 Data File : VN052608.D
 Acq On : 30 Nov 2018 15:48
 Operator : MD\SY
 Sample : PB115186ZHE#06
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB115186ZHE#06

Quant Time: Dec 01 02:18:35 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 28 01:11:38 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1204639	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1876883	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1659048	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	648944	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	645005	49.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.36%	
35) Dibromofluoromethane	7.59	113	572530	49.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.96%	
50) Toluene-d8	10.09	98	2280938	50.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.28%	
62) 4-Bromofluorobenzene	12.40	95	729991	48.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.46%	
Target Compounds						
16) Acetone	3.82	43	23770	6.871	ug/l	98
20) Methylene Chloride	4.56	84	20423	1.509	ug/l	94
43) Isopropyl Acetate	8.17	43	67864	3.707	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN113018\
Data File : VN052608.D
Acq On : 30 Nov 2018 15:48
Operator : MD\SY
Sample : PB115186ZHE#06
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB115186ZHE#06

Quant Time: Dec 01 02:18:35 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 28 01:11:38 2018
Response via : Initial Calibration

