

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN113018\
 Data File : VN052614.D
 Acq On : 30 Nov 2018 18:28
 Operator : MD\SY
 Sample : J5761-04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 AU-05-112918

Quant Time: Dec 01 02:32:41 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 28 01:11:38 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1213195	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1879377	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1674038	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	657293	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	646981	48.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.96%	
35) Dibromofluoromethane	7.59	113	574670	49.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.20%	
50) Toluene-d8	10.09	98	2293910	50.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.72%	
62) 4-Bromofluorobenzene	12.40	95	735574	48.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.08%	
Target Compounds						
16) Acetone	3.82	43	22906	6.574	ug/l	98
20) Methylene Chloride	4.55	84	9276520	680.378	ug/l	100
43) Isopropyl Acetate	8.17	43	71431	3.897	ug/l #	75
95) Naphthalene	15.14	128	20602	4.204	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN113018\
Data File : VN052614.D
Acq On : 30 Nov 2018 18:28
Operator : MD\SY
Sample : J5761-04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
AU-05-112918

Quant Time: Dec 01 02:32:41 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 28 01:11:38 2018
Response via : Initial Calibration

