

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN113020\
 Data File : VN064863.D
 Acq On : 30 Nov 2020 11:06
 Operator : JC/MD
 Sample : VN1130MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1130MBL01

Quant Time: Dec 01 01:45:43 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	163422	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	281795	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	289806	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	126756	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		7.99	65	97884	51.52	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	103.04%	
35) Dibromofluoromethane		7.55	113	89810	53.26	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	106.52%	
50) Toluene-d8		10.06	98	352030	51.10	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.20%	
62) 4-Bromofluorobenzene		12.37	95	132246	53.64	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	107.28%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN113020\
Data File : VN064863.D
Acq On : 30 Nov 2020 11:06
Operator : JC/MD
Sample : VN1130MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1130MBL01

Quant Time: Dec 01 01:45:43 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

