

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN113021\
 Data File : VN069682.D
 Acq On : 30 Nov 2021 13:04
 Operator : JC/MD
 Sample : M4848-01
 Misc : 4.71g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PSC124055

Quant Time: Dec 01 01:49:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 09 18:18:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	1447495	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	2504454	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2452441	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1057864	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.431	65	1070601	51.010	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	102.020%
35) Dibromofluoromethane	8.018	113	709672	47.246	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.500%
50) Toluene-d8	10.440	98	3170017	53.660	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.320%
62) 4-Bromofluorobenzene	12.731	95	1179894	54.367	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	108.740%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN113021\
Data File : VN069682.D
Acq On : 30 Nov 2021 13:04
Operator : JC/MD
Sample : M4848-01
Misc : 4.71g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PSC124055

Quant Time: Dec 01 01:49:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 09 18:18:22 2021
Response via : Initial Calibration

