

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN113021\
 Data File : VN069685.D
 Acq On : 30 Nov 2021 17:35
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 01 01:50:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 09 18:18:22 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	102	0.00
2 T	Dichlorodifluoromethane	0.534	0.456	14.6	83	0.00
3 P	Chloromethane	0.619	0.525	15.2	90	0.00
4 C	Vinyl Chloride	0.620	0.541	12.7#	88	0.00
5 T	Bromomethane	0.403	0.380	5.7	103	0.00
6 T	Chloroethane	0.419	0.345	17.7	85	0.00
7 T	Trichlorofluoromethane	0.959	0.804	16.2	85	0.00
8 T	Diethyl Ether	0.310	0.327	-5.5	110	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.452	0.477	-5.5	108	0.00
10 T	Methyl Iodide	0.554	0.612	-10.5	112	0.00
11 T	Tert butyl alcohol	0.131	0.117	10.7	92	-0.01
12 CM	1,1-Dichloroethene	0.405	0.413	-2.0#	103	0.00
13 T	Acrolein	0.021	0.021	0.0	118	0.00
14 T	Allyl chloride	0.896	0.973	-8.6	110	0.00
15 T	Acrylonitrile	0.335	0.354	-5.7	106	0.00
16 T	Acetone	0.391	0.362	7.4	94	0.00
17 T	Carbon Disulfide	0.958	0.753	21.4	80	0.00
18 T	Methyl Acetate	1.272	1.298	-2.0	107	0.00
19 T	Methyl tert-butyl Ether	1.715	1.921	-12.0	113	0.00
20 T	Methylene Chloride	0.555	0.549	1.1	107	0.00
21 T	trans-1,2-Dichloroethene	0.453	0.447	1.3	101	0.00
22 T	Diisopropyl ether	1.930	2.193	-13.6	114	0.00
23 T	Vinyl Acetate	1.493	1.679	-12.5	110	0.00
24 P	1,1-Dichloroethane	0.983	1.083	-10.2	113	0.00
25 T	2-Butanone	0.530	0.524	1.1	98	0.00
26 T	2,2-Dichloropropane	0.769	0.864	-12.4	110	0.00
27 T	cis-1,2-Dichloroethene	0.561	0.608	-8.4	109	0.00
28 T	Bromochloromethane	0.503	0.493	2.0	103	0.00
29 T	Tetrahydrofuran	0.325	0.327	-0.6	98	0.00
30 C	Chloroform	1.008	1.122	-11.3#	113	0.00
31 T	Cyclohexane	0.990	0.889	10.2	97	0.00
32 T	1,1,1-Trichloroethane	0.841	0.927	-10.2	108	0.00
33 S	1,2-Dichloroethane-d4	0.725	0.738	-1.8	110	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	101	0.00
35 S	Dibromofluoromethane	0.300	0.317	-5.7	109	0.00
36 T	1,1-Dichloropropene	0.421	0.454	-7.8	103	0.00
37 T	Ethyl Acetate	0.569	0.629	-10.5	103	0.00
38 T	Carbon Tetrachloride	0.406	0.459	-13.1	107	0.00
39 T	Methylcyclohexane	0.523	0.528	-1.0	94	0.00
40 TM	Benzene	1.299	1.403	-8.0	105	0.00
41 T	Methacrylonitrile	0.294	0.328	-11.6	113	0.00
42 TM	1,2-Dichloroethane	0.508	0.589	-15.9	113	0.00
43 T	Isopropyl Acetate	0.913	1.031	-12.9	107	0.00
44 TM	Trichloroethene	0.311	0.347	-11.6	109	0.00
45 C	1,2-Dichloropropane	0.347	0.409	-17.9#	113	0.00
46 T	Dibromomethane	0.232	0.258	-11.2	107	0.00
47 T	Bromodichloromethane	0.434	0.530	-22.1	113	0.00
48 T	Methyl methacrylate	0.430	0.471	-9.5	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN113021\
 Data File : VN069685.D
 Acq On : 30 Nov 2021 17:35
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 01 01:50:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 09 18:18:22 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.008	0.0	94	0.00
50 S	Toluene-d8	1.179	1.204	-2.1	102	0.00
51 T	4-Methyl-2-Pentanone	0.590	0.665	-12.7	103	0.00
52 CM	Toluene	0.809	0.892	-10.3#	104	0.00
53 T	t-1,3-Dichloropropene	0.453	0.571	-26.0#	113	0.00
54 T	cis-1,3-Dichloropropene	0.495	0.614	-24.0	112	0.00
55 T	1,1,2-Trichloroethane	0.326	0.387	-18.7	112	0.00
56 T	Ethyl methacrylate	0.540	0.635	-17.6	108	0.00
57 T	1,3-Dichloropropane	0.573	0.663	-15.7	111	0.00
58 T	2-Chloroethyl Vinyl ether	0.125	0.146	-16.8	108	0.00
59 T	2-Hexanone	0.439	0.461	-5.0	95	0.00
60 T	Dibromochloromethane	0.305	0.389	-27.5#	114	0.00
61 T	1,2-Dibromoethane	0.330	0.388	-17.6	108	0.00
62 S	4-Bromofluorobenzene	0.433	0.476	-9.9	110	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	99	0.00
64 T	Tetrachloroethene	0.302	0.331	-9.6	106	0.00
65 PM	Chlorobenzene	0.931	1.054	-13.2	109	0.00
66 T	1,1,1,2-Tetrachloroethane	0.323	0.391	-21.1	112	0.00
67 C	Ethyl Benzene	1.716	1.922	-12.0#	106	0.00
68 T	m/p-Xylenes	0.631	0.704	-11.6	104	0.00
69 T	o-Xylene	0.629	0.717	-14.0	106	0.00
70 T	Styrene	1.011	1.187	-17.4	108	0.00
71 P	Bromoform	0.230	0.293	-27.4#	112	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
73 T	Isopropylbenzene	3.981	4.547	-14.2	108	0.00
74 T	N-amyl acetate	1.687	1.835	-8.8	98	0.00
75 P	1,1,2,2-Tetrachloroethane	1.343	1.428	-6.3	102	0.00
76 T	1,2,3-Trichloropropane	1.154	1.202	-4.2	98	0.00
77 T	Bromobenzene	0.910	1.029	-13.1	108	0.00
78 T	n-propylbenzene	4.480	5.185	-15.7	107	0.00
79 T	2-Chlorotoluene	2.774	3.172	-14.3	109	0.00
80 T	1,3,5-Trimethylbenzene	3.218	3.688	-14.6	107	0.00
81 T	trans-1,4-Dichloro-2-butene	0.344	0.381	-10.8	101	0.00
82 T	4-Chlorotoluene	2.644	3.034	-14.8	109	0.00
83 T	tert-Butylbenzene	2.820	3.316	-17.6	109	0.00
84 T	1,2,4-Trimethylbenzene	3.107	3.591	-15.6	106	0.00
85 T	sec-Butylbenzene	3.940	4.602	-16.8	107	0.00
86 T	p-Isopropyltoluene	3.128	3.674	-17.5	106	0.00
87 T	1,3-Dichlorobenzene	1.579	1.755	-11.1	103	0.00
88 T	1,4-Dichlorobenzene	1.549	1.705	-10.1	103	0.00
89 T	n-Butylbenzene	2.608	3.011	-15.5	103	0.00
90 T	Hexachloroethane	0.513	0.629	-22.6	108	0.00
91 T	1,2-Dichlorobenzene	1.502	1.711	-13.9	105	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.217	0.250	-15.2	99	0.00
93 T	1,2,4-Trichlorobenzene	0.662	0.790	-19.3	100	0.00
94 T	Hexachlorobutadiene	0.390	0.486	-24.6	111	0.00
95 T	Naphthalene	1.775	1.940	-9.3	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN113021\
Data File : VN069685.D
Acq On : 30 Nov 2021 17:35
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Dec 01 01:50:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 09 18:18:22 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.634	0.746	-17.7	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6