

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN113022\
 Data File : VN075495.D
 Acq On : 30 Nov 2022 11:52
 Operator : JC\MD
 Sample : N5795-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-12

Quant Time: Dec 01 02:14:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	130147	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	229370	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	206434	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	80751	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	34201	48.956	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.920%
35) Dibromofluoromethane	8.171	113	42018	46.633	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.260%
50) Toluene-d8	10.571	98	91085	44.883	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.760%
62) 4-Bromofluorobenzene	12.847	95	60982	40.711	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	81.420%#
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	13852	21.257	ug/l #	86
20) Methylene Chloride	5.295	84	6675	3.872	ug/l	97
37) Ethyl Acetate	7.571	43	16459	9.303	ug/l	98
43) Isopropyl Acetate	8.688	43	77693	27.357	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN113022\
Data File : VN075495.D
Acq On : 30 Nov 2022 11:52
Operator : JC\MD
Sample : N5795-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-12

Quant Time: Dec 01 02:14:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

