

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN113022\
 Data File : VN075496.D
 Acq On : 30 Nov 2022 13:04
 Operator : JC\MD
 Sample : PB149341TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB149341TB

Quant Time: Dec 01 02:14:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	139193	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	239455	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	219473	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	91289	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	37713	50.475	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 100.940%	
35) Dibromofluoromethane	8.171	113	45436	48.302	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 96.600%	
50) Toluene-d8	10.571	98	104466	49.308	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 98.620%	
62) 4-Bromofluorobenzene	12.853	95	75960	48.575	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 97.140%	
Target Compounds						
						Qvalue
16) Acetone	4.448	43	11501	16.502	ug/l #	81
37) Ethyl Acetate	7.571	43	20884	11.306	ug/l	98
43) Isopropyl Acetate	8.688	43	101250	34.150	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN113022\
Data File : VN075496.D
Acq On : 30 Nov 2022 13:04
Operator : JC\MD
Sample : PB149341TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB149341TB

Quant Time: Dec 01 02:14:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

