

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN113022\
 Data File : VN075512.D
 Acq On : 30 Nov 2022 19:27
 Operator : JC\MD
 Sample : N5809-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 HOWE-DRUM-A-1129

Manual Integrations
 APPROVED

Reviewed By : John Carlone 12/01/2022
 Supervised By : Mahesh Dadoda 12/01/2022

Quant Time: Dec 01 02:19:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	132451	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	237952	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	218430	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	85148	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	35263	49.598	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	99.200%		
35) Dibromofluoromethane	8.177	113	43050	46.055	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	92.100%		
50) Toluene-d8	10.571	98	92374	43.876	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	87.760%		
62) 4-Bromofluorobenzene	12.853	95	67173	43.227	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	86.460%		
Target Compounds					Qvalue	
16) Acetone	4.448	43	15721	23.706	ug/l	98
18) Methyl Acetate	5.042	43	5640	2.134	ug/l	100
20) Methylene Chloride	5.289	84	5180	2.641	ug/l #	77
37) Ethyl Acetate	7.571	43	22999	12.530	ug/l	98
43) Isopropyl Acetate	8.694	43	183342m	62.230	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN113022\
Data File : VN075512.D
Acq On : 30 Nov 2022 19:27
Operator : JC\MD
Sample : N5809-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
HOWE-DRUM-A-1129

Quant Time: Dec 01 02:19:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 12/01/2022
Supervised By :Mahesh Dadoda 12/01/2022

