

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120120\
 Data File : VN064880.D
 Acq On : 1 Dec 2020 14:16
 Operator : JC/MD
 Sample : L4908-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Dec 02 00:32:15 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	157782	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	287414	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	291588	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	129018	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	99580	54.29	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	108.58%		
35) Dibromofluoromethane	7.55	113	90620	52.69	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	105.38%		
50) Toluene-d8	10.06	98	351552	50.03	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	100.06%		
62) 4-Bromofluorobenzene	12.38	95	133708	53.17	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	106.34%		

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120120\
Data File : VN064880.D
Acq On : 1 Dec 2020 14:16
Operator : JC/MD
Sample : L4908-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Dec 02 00:32:15 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

