

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120120\
 Data File : VN064881.D
 Acq On : 1 Dec 2020 14:40
 Operator : JC/MD
 Sample : L4908-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANA

Quant Time: Dec 02 00:33:15 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	154135	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	278426	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	283238	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	126218	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		7.99	65	97296	54.30	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	108.60%	
35) Dibromofluoromethane		7.55	113	88855	53.33	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	106.66%	
50) Toluene-d8		10.06	98	342709	50.35	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	100.70%	
62) 4-Bromofluorobenzene		12.38	95	131690	54.06	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	108.12%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120120\
Data File : VN064881.D
Acq On : 1 Dec 2020 14:40
Operator : JC/MD
Sample : L4908-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANA

Quant Time: Dec 02 00:33:15 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

