

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120121\
 Data File : VN069693.D
 Acq On : 1 Dec 2021 16:09
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Dec 02 07:26:10 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 09 18:18:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1447276	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2536496	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	2600622	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	996464	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	1081583	51.541	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	103.080%
35) Dibromofluoromethane	8.021	113	743182	48.852	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.700%
50) Toluene-d8	10.443	98	3227764	53.947	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.900%
62) 4-Bromofluorobenzene	12.731	95	1197040	54.460	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	108.920%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120121\
Data File : VN069693.D
Acq On : 1 Dec 2021 16:09
Operator : JC/MD
Sample : IBLK
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Dec 02 07:26:10 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 09 18:18:22 2021
Response via : Initial Calibration

