

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120121\
 Data File : VN069694.D
 Acq On : 1 Dec 2021 16:34
 Operator : JC/MD
 Sample : M4892-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EFF-1221

Quant Time: Dec 02 07:26:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 09 18:18:22 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1444806	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2550738	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2585642	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	990457	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	1091031	52.081	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 104.160%	
35) Dibromofluoromethane	8.021	113	753254	49.238	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 98.480%	
50) Toluene-d8	10.443	98	3190969	53.035	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 106.060%	
62) 4-Bromofluorobenzene	12.731	95	1197810	54.191	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 108.380%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	5.377	59	91897	24.344	ug/l	100
16) Acetone	4.291	43	36787	3.255	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120121\
Data File : VN069694.D
Acq On : 1 Dec 2021 16:34
Operator : JC/MD
Sample : M4892-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EFF-1221

Quant Time: Dec 02 07:26:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 09 18:18:22 2021
Response via : Initial Calibration

