

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120121\
 Data File : VN069695.D
 Acq On : 1 Dec 2021 17:00
 Operator : JC/MD
 Sample : M4892-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MID-1221

Quant Time: Dec 02 07:26:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 09 18:18:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1490827	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2591504	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	2602725	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	984913	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	1098329	50.810	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 101.620%	
35) Dibromofluoromethane	8.024	113	761077	48.967	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 97.940%	
50) Toluene-d8	10.443	98	3273285	53.547	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 107.100%	
62) 4-Bromofluorobenzene	12.733	95	1190303	53.004	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 106.000%	
Target Compounds						Qvalue
11) Tert butyl alcohol	5.377	59	81510	20.926	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120121\
Data File : VN069695.D
Acq On : 1 Dec 2021 17:00
Operator : JC/MD
Sample : M4892-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MID-1221

Quant Time: Dec 02 07:26:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 09 18:18:22 2021
Response via : Initial Calibration

