

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120121\
 Data File : VN069696.D
 Acq On : 1 Dec 2021 17:25
 Operator : JC/MD
 Sample : M4892-03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 INF-1221

Quant Time: Dec 10 00:17:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 09 18:18:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1639341	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2795862	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2931565	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1404979	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	1158900	48.76	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	97.52%		
35) Dibromofluoromethane	8.024	113	832742	49.66	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	99.32%		
50) Toluene-d8	10.440	98	3607279	54.70	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	109.40%		
62) 4-Bromofluorobenzene	12.731	95	1509062	62.29	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	124.58%		
Target Compounds					Qvalue	
11) Tert butyl alcohol	5.380	59	83444	19.48	ug/l #	71
25) 2-Butanone	7.345	43	45553	2.62	ug/l	90
31) Cyclohexane	8.096	56	1254487	38.63	ug/l	90
39) Methylcyclohexane	9.462	83	529985	18.11	ug/l	89
40) Benzene	8.461	78	19043043	262.25	ug/l	96
52) Toluene	10.505	92	2375935	52.52	ug/l	97
67) Ethyl Benzene	11.843	91	14634744	145.42	ug/l	98
68) m/p-Xylenes	11.956	106	12812683	346.16	ug/l #	1
69) o-Xylene	12.280	106	2402040	65.10	ug/l	92
73) Isopropylbenzene	12.578	105	1541187	13.78	ug/l	98
78) n-propylbenzene	12.919	91	3184764	25.30	ug/l	95
80) 1,3,5-Trimethylbenzene	13.058	105	1872492	20.71	ug/l	96
84) 1,2,4-Trimethylbenzene	13.369	105	12480322	142.95	ug/l	99
86) p-Isopropyltoluene	13.616	119	55812	0.64	ug/l #	78
95) Naphthalene	15.504	128	2669970	43.45	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120121\
Data File : VN069696.D
Acq On : 1 Dec 2021 17:25
Operator : JC/MD
Sample : M4892-03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
INF-1221

Quant Time: Dec 10 00:17:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 09 18:18:22 2021
Response via : Initial Calibration

