

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120122\
 Data File : VN075516.D
 Acq On : 01 Dec 2022 15:35
 Operator : JC\MD
 Sample : VN1201MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1201MBL01

Quant Time: Dec 02 04:36:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	146333	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	239127	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	208687	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	84448	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	38341	48.811	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.620%
35) Dibromofluoromethane	8.171	113	48145	51.252	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.500%
50) Toluene-d8	10.571	98	96958	45.827	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.660%
62) 4-Bromofluorobenzene	12.853	95	68543	43.892	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.780%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120122\
Data File : VN075516.D
Acq On : 01 Dec 2022 15:35
Operator : JC\MD
Sample : VN1201MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1201MBL01

Quant Time: Dec 02 04:36:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

