

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120122\
 Data File : VN075520.D
 Acq On : 01 Dec 2022 17:30
 Operator : JC\MD
 Sample : N5787-01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-24-STOCK

Quant Time: Dec 02 04:38:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	229483	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	373024	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	312440	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	134297	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	64019	51.971	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 103.940%	
35) Dibromofluoromethane	8.171	113	78851	53.810	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 107.620%	
50) Toluene-d8	10.565	98	173260	52.497	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 105.000%	
62) 4-Bromofluorobenzene	12.853	95	118380	48.595	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 97.180%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120122\
Data File : VN075520.D
Acq On : 01 Dec 2022 17:30
Operator : JC\MD
Sample : N5787-01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SB-24-STOCK

Quant Time: Dec 02 04:38:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

