

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120123\
 Data File : VN080321.D
 Acq On : 01 Dec 2023 15:12
 Operator : JC\MD
 Sample : 05607-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Dec 01 22:34:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	171124	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	309537	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	319008	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	144474	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	149364	52.862	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.720%
35) Dibromofluoromethane	8.171	113	118371	51.827	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.660%
50) Toluene-d8	10.565	98	404711	49.156	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.320%
62) 4-Bromofluorobenzene	12.853	95	154798	49.777	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.560%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120123\
Data File : VN080321.D
Acq On : 01 Dec 2023 15:12
Operator : JC\MD
Sample : 05607-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Dec 01 22:34:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

