


cis-1,3-Dichloropropene

Response Ratio



$$\text{Response} = 6.380\text{e-}001 * \text{Amt} - 2.550\text{e-}002$$


Coef of Det (r^2) = 0.998280 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N120221W.M

Calibration Table Last Updated: Fri Dec 03 08:46:47 2021

t-1,3-Dichloropropene

Response Ratio



$$\text{Response} = 6.181\text{e-}001 * \text{Amt} - 3.454\text{e-}002$$


Coef of Det (r^2) = 0.997912 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N120221W.M

Calibration Table Last Updated: Fri Dec 03 08:46:47 2021

Ethyl methacrylate

Response Ratio



$$\text{Response} = 7.085\text{e-}001 * \text{Amt} - 3.230\text{e-}002$$


Coef of Det (r^2) = 0.998255 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N120221W.M

Calibration Table Last Updated: Fri Dec 03 08:46:47 2021

Ethyl methacrylate

Response Ratio



$$\text{Response} = 7.085\text{e-}001 * \text{Amt} - 3.230\text{e-}002$$


Coef of Det (r^2) = 0.998255 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N120221W.M

Calibration Table Last Updated: Fri Dec 03 08:46:47 2021

2-Chloroethyl Vinyl ether


Response Ratio



Response = $1.971 \times 10^{-1} \times \text{Amt} - 7.249 \times 10^{-2}$
Coef of Det (r^2) = 0.996039 Curve Fit: Linear
Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N120221W.M
Calibration Table Last Updated: Fri Dec 03 08:46:47 2021

Dibromochloromethane

Response Ratio



$$\text{Response} = 4.032\text{e-}001 * \text{Amt} - 1.916\text{e-}002$$


Coef of Det (r^2) = 0.998370 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N120221W.M

Calibration Table Last Updated: Fri Dec 03 08:46:47 2021

Bromoform


Response Ratio



$\text{Response} = 3.230\text{e-}001 * \text{Amt} - 1.764\text{e-}002$
 Coef of Det (r^2) = 0.998365 Curve Fit: Linear
 Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N120221W.M
 Calibration Table Last Updated: Fri Dec 03 08:46:47 2021

trans-1,4-Dichloro-2-butene

Response Ratio



$$\text{Response} = 4.768\text{e-}001 * \text{Amt} - 3.873\text{e-}002$$


Coef of Det (r^2) = 0.998440 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N120221W.M

Calibration Table Last Updated: Fri Dec 03 08:46:47 2021

1,2,4-Trichlorobenzene

Response Ratio



$$\text{Response} = 8.704\text{e-}001 * \text{Amt} - 5.010\text{e-}002$$


Coef of Det (r^2) = 0.998076 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N120221W.M

Calibration Table Last Updated: Fri Dec 03 08:46:47 2021

Naphthalene

Response Ratio



$$\text{Response} = 2.655\text{e}+000 * \text{Amt} - 2.116\text{e}-001$$


Coef of Det (r^2) = 0.996782 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N120221W.M

Calibration Table Last Updated: Fri Dec 03 08:46:47 2021

1,2,3-Trichlorobenzene

Response Ratio



$$\text{Response} = 8.404\text{e-}001 * \text{Amt} - 4.094\text{e-}002$$

Coef of Det (r^2) = 0.998252 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N120221W.M

Calibration Table Last Updated: Fri Dec 03 08:46:47 2021