

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120222\
 Data File : VN075525.D
 Acq On : 02 Dec 2022 12:21
 Operator : JC\MD
 Sample : VN1202MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1202MBL01

Quant Time: Dec 03 00:30:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	125628	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	221154	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	186535	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	73602	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	35304	52.352	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.700%
35) Dibromofluoromethane	8.171	113	43947	50.586	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.180%
50) Toluene-d8	10.571	98	86719	44.319	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	88.640%
62) 4-Bromofluorobenzene	12.853	95	57070	39.515	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	79.040%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120222\
Data File : VN075525.D
Acq On : 02 Dec 2022 12:21
Operator : JC\MD
Sample : VN1202MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1202MBL01

Quant Time: Dec 03 00:30:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

