

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120222\
 Data File : VN075530.D
 Acq On : 02 Dec 2022 14:29
 Operator : JC\MD
 Sample : PB149377TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB149377TB

Quant Time: Dec 03 00:32:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	147488	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	9.106	114	245541	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	224530	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	97130	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	39423	49.796	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.600%
35) Dibromofluoromethane	8.177	113	50445	52.298	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.600%
50) Toluene-d8	10.571	98	108720	50.044	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.080%
62) 4-Bromofluorobenzene	12.847	95	77376	48.254	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.500%
Target Compounds						
						Qvalue
16) Acetone	4.453	43	12908	17.480	ug/l	# 88
20) Methylene Chloride	5.295	84	4292	1.629	ug/l	# 80
37) Ethyl Acetate	7.571	43	22269	11.757	ug/l	# 93
43) Isopropyl Acetate	8.694	43	100022	32.900	ug/l	# 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120222\
Data File : VN075530.D
Acq On : 02 Dec 2022 14:29
Operator : JC\MD
Sample : PB149377TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB149377TB

Quant Time: Dec 03 00:32:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

