

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120222\
 Data File : VN075531.D
 Acq On : 02 Dec 2022 14:53
 Operator : JC\MD
 Sample : N5793-02 10X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 B-40-SB01

Quant Time: Dec 03 00:32:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	147368	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	248555	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	220174	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	90721	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	46608	58.919	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	117.840%
35) Dibromofluoromethane	8.171	113	56701	58.071	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	116.140%
50) Toluene-d8	10.571	98	121815	55.392	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	110.780%
62) 4-Bromofluorobenzene	12.853	95	83718	51.576	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.160%
Target Compounds						
16) Acetone	4.453	43	2955	4.005	ug/l	94
43) Isopropyl Acetate	8.694	43	9394	3.052	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120222\
Data File : VN075531.D
Acq On : 02 Dec 2022 14:53
Operator : JC\MD
Sample : N5793-02 10X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
B-40-SB01

Quant Time: Dec 03 00:32:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

