

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120222\
 Data File : VN075532.D
 Acq On : 02 Dec 2022 15:17
 Operator : JC\MD
 Sample : N5833-02 10X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 COMPOSITE-1

Quant Time: Dec 03 00:33:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	136482	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	230926	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	205986	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	85054	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	41886	57.173	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.340%
35) Dibromofluoromethane	8.171	113	51802	57.104	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	114.200%
50) Toluene-d8	10.571	98	103796	50.802	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.600%
62) 4-Bromofluorobenzene	12.853	95	74397	49.332	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.660%
Target Compounds						
43) Isopropyl Acetate	8.700	43	7115	2.488	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120222\
Data File : VN075532.D
Acq On : 02 Dec 2022 15:17
Operator : JC\MD
Sample : N5833-02 10X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
COMPOSITE-1

Quant Time: Dec 03 00:33:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

