

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120222\
 Data File : VN075541.D
 Acq On : 02 Dec 2022 18:53
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Dec 03 00:37:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	151905	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	248603	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	214394	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	91104	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	40859	50.109	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	100.220%
35) Dibromofluoromethane	8.171	113	52306	53.560	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.120%
50) Toluene-d8	10.565	98	105012	47.742	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.480%
62) 4-Bromofluorobenzene	12.853	95	73974	45.564	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.120%
Target Compounds						
52) Toluene	10.630	92	3559	0.847	ug/l	98
68) m/p-Xylenes	12.065	106	1428	0.466	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120222\
Data File : VN075541.D
Acq On : 02 Dec 2022 18:53
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Dec 03 00:37:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

