

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120222\
 Data File : VN075523.D
 Acq On : 02 Dec 2022 11:23
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 03 00:23:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	85	0.00
2 T	Dichlorodifluoromethane	0.382	0.395	-3.4	89	0.00
3 P	Chloromethane	0.476	0.484	-1.7	94	0.00
4 C	Vinyl Chloride	0.600	0.582	3.0#	85	0.00
5 T	Bromomethane	0.527	0.592	-12.3	91	0.02
6 T	Chloroethane	0.454	0.450	0.9	90	0.01
7 T	Trichlorofluoromethane	0.857	0.853	0.5	87	0.00
8 T	Diethyl Ether	0.300	0.340	-13.3	93	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.473	0.497	-5.1	91	0.00
10 T	Methyl Iodide	0.766	0.852	-11.2	95	0.00
11 T	Tert butyl alcohol	0.113	0.094	16.8	78	-0.01
12 CM	1,1-Dichloroethene	0.463	0.473	-2.2#	88	0.00
13 T	Acrolein	0.098	0.100	-2.0	91	0.00
14 T	Allyl chloride	0.599	0.642	-7.2	99	0.00
15 T	Acrylonitrile	0.258	0.274	-6.2	92	0.01
16 T	Acetone	0.250	0.242	3.2	89	0.00
17 T	Carbon Disulfide	1.203	1.175	2.3	84	0.00
18 T	Methyl Acetate	0.741	0.702	5.3	94	0.00
19 T	Methyl tert-butyl Ether	1.640	1.790	-9.1	96	0.00
20 T	Methylene Chloride	0.587	0.590	-0.5	97	0.00
21 T	trans-1,2-Dichloroethene	0.528	0.538	-1.9	90	0.00
22 T	Diisopropyl ether	1.367	1.500	-9.7	93	0.00
23 T	Vinyl Acetate	1.087	1.091	-0.4	83	0.00
24 P	1,1-Dichloroethane	0.877	0.955	-8.9	93	0.00
25 T	2-Butanone	0.350	0.349	0.3	90	0.00
26 T	2,2-Dichloropropane	0.815	0.817	-0.2	86	0.00
27 T	cis-1,2-Dichloroethene	0.597	0.677	-13.4	94	0.00
28 T	Bromochloromethane	0.356	0.363	-2.0	89	0.00
29 T	Tetrahydrofuran	0.213	0.225	-5.6	91	0.00
30 C	Chloroform	1.007	1.074	-6.7#	95	0.00
31 T	Cyclohexane	0.853	0.797	6.6	86	0.00
32 T	1,1,1-Trichloroethane	0.915	0.955	-4.4	89	0.00
33 S	1,2-Dichloroethane-d4	0.268	0.281	-4.9	90	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	84	0.00
35 S	Dibromofluoromethane	0.196	0.218	-11.2	89	0.00
36 T	1,1-Dichloropropene	0.428	0.454	-6.1	88	0.00
37 T	Ethyl Acetate	0.386	0.429	-11.1	89	0.00
38 T	Carbon Tetrachloride	0.471	0.525	-11.5	90	0.00
39 T	Methylcyclohexane	0.536	0.576	-7.5	87	0.00
40 TM	Benzene	1.311	1.465	-11.7	94	0.00
41 T	Methacrylonitrile	0.197	0.216	-9.6	93	0.00
42 TM	1,2-Dichloroethane	0.439	0.506	-15.3	98	0.00
43 T	Isopropyl Acetate	0.619	0.703	-13.6	93	0.00
44 TM	Trichloroethene	0.359	0.388	-8.1	91	0.00
45 C	1,2-Dichloropropane	0.313	0.348	-11.2#	93	0.00
46 T	Dibromomethane	0.236	0.276	-16.9	98	0.00
47 T	Bromodichloromethane	0.458	0.527	-15.1	95	0.00
48 T	Methyl methacrylate	0.286	0.325	-13.6	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120222\
 Data File : VN075523.D
 Acq On : 02 Dec 2022 11:23
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 03 00:23:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.009	-12.5	84	0.00
50 S	Toluene-d8	0.442	0.461	-4.3	83	0.00
51 T	4-Methyl-2-Pentanone	0.394	0.446	-13.2	93	0.00
52 CM	Toluene	0.845	0.982	-16.2#	94	0.00
53 T	t-1,3-Dichloropropene	0.460	0.558	-21.3	93	0.00
54 T	cis-1,3-Dichloropropene	0.499	0.601	-20.4	95	0.00
55 T	1,1,2-Trichloroethane	0.344	0.405	-17.7	100	0.00
56 T	Ethyl methacrylate	0.465	0.565	-21.5	93	0.00
57 T	1,3-Dichloropropane	0.561	0.649	-15.7	97	0.00
58 T	2-Chloroethyl Vinyl ether	0.168	0.187	-11.3	84	0.00
59 T	2-Hexanone	0.296	0.329	-11.1	90	0.00
60 T	Dibromochloromethane	0.353	0.456	-29.2#	97	0.00
61 T	1,2-Dibromoethane	0.357	0.425	-19.0	97	0.00
62 S	4-Bromofluorobenzene	0.327	0.365	-11.6	87	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	87	0.00
64 T	Tetrachloroethene	0.382	0.372	2.6	90	0.00
65 PM	Chlorobenzene	1.033	1.114	-7.8	94	0.00
66 T	1,1,1,2-Tetrachloroethane	0.385	0.435	-13.0	96	0.00
67 C	Ethyl Benzene	1.762	1.952	-10.8#	92	0.00
68 T	m/p-Xylenes	0.715	0.801	-12.0	95	0.00
69 T	o-Xylene	0.711	0.784	-10.3	96	0.00
70 T	Styrene	1.121	1.312	-17.0	95	0.00
71 P	Bromoform	0.293	0.349	-19.1	94	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	83	0.00
73 T	Isopropylbenzene	3.653	3.913	-7.1	91	0.00
74 T	N-amyl acetate	1.123	1.259	-12.1	91	0.00
75 P	1,1,2,2-Tetrachloroethane	1.114	1.239	-11.2	98	0.00
76 T	1,2,3-Trichloropropane	0.945	1.054	-11.5	99	0.00
77 T	Bromobenzene	0.902	0.995	-10.3	95	0.00
78 T	n-propylbenzene	4.077	4.449	-9.1	90	0.00
79 T	2-Chlorotoluene	2.456	2.750	-12.0	96	0.00
80 T	1,3,5-Trimethylbenzene	3.035	3.422	-12.8	92	0.00
81 T	trans-1,4-Dichloro-2-butene	0.320	0.382	-19.4	95	0.00
82 T	4-Chlorotoluene	2.456	2.750	-12.0	96	0.00
83 T	tert-Butylbenzene	2.767	2.972	-7.4	90	0.00
84 T	1,2,4-Trimethylbenzene	3.044	3.489	-14.6	94	0.00
85 T	sec-Butylbenzene	3.755	4.115	-9.6	88	0.00
86 T	p-Isopropyltoluene	3.159	3.621	-14.6	91	0.00
87 T	1,3-Dichlorobenzene	1.709	1.904	-11.4	94	0.00
88 T	1,4-Dichlorobenzene	1.736	1.908	-9.9	95	0.00
89 T	n-Butylbenzene	2.668	2.882	-8.0	86	0.00
90 T	Hexachloroethane	0.545	0.624	-14.5	90	0.00
91 T	1,2-Dichlorobenzene	1.666	1.887	-13.3	93	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.211	0.229	-8.5	87	0.00
93 T	1,2,4-Trichlorobenzene	0.854	0.976	-14.3	87	0.00
94 T	Hexachlorobutadiene	0.445	0.477	-7.2	84	0.00
95 T	Naphthalene	2.685	2.941	-9.5	85	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120222\
Data File : VN075523.D
Acq On : 02 Dec 2022 11:23
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Dec 03 00:23:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.839	0.932	-11.1	88	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6