

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120225\
 Data File : VN088398.D
 Acq On : 02 Dec 2025 10:40
 Operator : JC\MD
 Sample : VN1202MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1202MBL01

Quant Time: Dec 03 01:24:40 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112425W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 25 01:34:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	171979	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	394708	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	389937	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	179883	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	204008	62.970	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	125.940%#
35) Dibromofluoromethane	8.147	113	149305	54.584	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.160%
50) Toluene-d8	10.541	98	523434	51.431	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.860%
62) 4-Bromofluorobenzene	12.823	95	184788	52.916	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	105.840%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120225\
Data File : VN088398.D
Acq On : 02 Dec 2025 10:40
Operator : JC\MD
Sample : VN1202MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1202MBL01

Quant Time: Dec 03 01:24:40 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112425W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 25 01:34:52 2025
Response via : Initial Calibration

