

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120318\
 Data File : VN052621.D
 Acq On : 3 Dec 2018 10:27
 Operator : MD\SY
 Sample : PB115242TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 4 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB115242TB

Quant Time: Dec 04 01:00:18 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 03 03:24:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1252218	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1908222	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1646345	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	589351	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	660737	48.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.92%	
35) Dibromofluoromethane	7.59	113	587370	49.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.86%	
50) Toluene-d8	10.09	98	2343595	51.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.36%	
62) 4-Bromofluorobenzene	12.40	95	703961	45.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.50%	
Target Compounds						
16) Acetone	3.82	43	14852	4.130	ug/l	90
20) Methylene Chloride	4.55	84	175754	12.489	ug/l	99
43) Isopropyl Acetate	8.17	43	42227	2.269	ug/l #	88
95) Naphthalene	15.14	128	3317	3.568	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120318\
Data File : VN052621.D
Acq On : 3 Dec 2018 10:27
Operator : MD\SY
Sample : PB115242TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 4 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB115242TB

Quant Time: Dec 04 01:00:18 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 03 03:24:02 2018
Response via : Initial Calibration

