

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120318\
 Data File : VN052637.D
 Acq On : 3 Dec 2018 17:49
 Operator : MD\SY
 Sample : PB115231ZHE#08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB115231ZHE#08

Quant Time: Dec 04 01:25:13 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 03 03:24:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1234667	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1909813	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1663064	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	595841	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	646475	48.09	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.18%	
35) Dibromofluoromethane	7.59	113	585383	49.72	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.44%	
50) Toluene-d8	10.09	98	2348470	51.24	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.48%	
62) 4-Bromofluorobenzene	12.40	95	721415	46.84	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.68%	
Target Compounds						
16) Acetone	3.82	43	15059	4.247	ug/l	99
20) Methylene Chloride	4.56	84	91799	6.616	ug/l	98
43) Isopropyl Acetate	8.17	43	65363	3.509	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120318\
Data File : VN052637.D
Acq On : 3 Dec 2018 17:49
Operator : MD\SY
Sample : PB115231ZHE#08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB115231ZHE#08

Quant Time: Dec 04 01:25:13 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 03 03:24:02 2018
Response via : Initial Calibration

